

Material declaration concerning natural rubber latex

Based on available material data, VBM Medizintechnik GmbH confirms, that our medical devices with all their containing parts and components, are manufactured without the use of natural rubber latex. Excluded are the below mentioned articles, which are labeled with the symbol "LATEX".

**Warning:**

Products with natural rubber latex can cause allergic skin reactions and allergic asthma-like symptoms!

Product	REF	Component which contains natural rubber latex
Pressure infusion cuffs with spare part rubber bladder	52-01-300	Rubber bladder inc. tubings
	52-08-300	
	53-01-150	
	53-08-150	
	52-82-500	
	51-81-050	
Guedel airway with elastic necktape, size 0	40-26-009-1	Elastic tape
	40-06-040	
Endotape oral/nasal	40-8018-044	Elastic tape

This list only considers products from our current portfolio. If you need information about products which are not included in the current product catalogue on our homepage, please contact VBM.